

# EC - DECLARATION OF CONFORMITY **04**

We,

Dwyer Instruments, Inc.  
102 Indiana Highway 212  
Michigan City, IN 46360 USA  
+1-219-879-8868

declare under our sole responsibility that our Series 616KD-A and 616KD-B to which this declaration relates are in conformity with the following Directives and harmonized standards:

Directive 2011/65/EU Restriction of Certain Hazardous Substances in Electrical and Electronic Equipment

Directive 2014/30/EU Electromagnetic Compatibility (EMC)

EN 61326-1:2013 Electrical Equipment for Measurement, Control and Laboratory Use – EMC Requirements

IEC 61000-4-3:2006+A1:2007+A2:2010 Part 4-3 Testing and Measurement Techniques – Radiated, Radio-Frequency, Electromagnetic Field Immunity Test

IEC 61000-4-4:2012 Part 4-4 Testing and Measurement Techniques – Electrical Fast Transient/Burst Immunity Test

IEC 61000-4-6:2013 Part 4-6 Testing and Measurement Techniques – Immunity to Conducted Disturbances, Induced by Radio-Frequency Fields

EN 55011:2009+A1:2010 Industrial, Scientific, and Medical (ISM) Radio Frequency Equipment – Electromagnetic Disturbance Characteristics – Limits and Methods of Measurement

Models included:

616KD-A-XX and 616KD-A-XX-V where XX may be 02, 03, 04, 05, 06, 07, 08, 12, 13, or 14.

616KD-B-XX and 616KD-B-XX-V where XX may be 00, 01, 02, 03, 04, 05, 06, 07, 08, 10, 11, 12, 13, or 14.

The authorized representative located within the Community is:

Dwyer Instruments Ltd  
Unit 16 The Wye Estate, London Road  
High Wycombe, HP11 1LH-U.K.  
+44 (0) 1494 461707

On behalf of Dwyer Instruments, Inc.,



Doug McCall  
Authorized Signatory  
Senior Regulatory Engineer  
Title of Authorized Signatory  
October 01, 2015  
Date of Signature

Dwyer Instruments, Inc.  
102 Indiana Highway 212  
Michigan City, IN 46360 USA  
+1-219-879-8868