



10

## EC - DECLARATION OF CONFORMITY

We,

Dwyer Instruments, Inc.  
P.O. Box 373  
Michigan City, IN 46361, USA  
(219) 879-8868

declare under our sole responsibility that our Model 450A Digital Pocket Size Carbon Monoxide Meter to which this declaration relates are in conformity with the following EC Directives and harmonized standards:

Directive 104/108/EC (EMC)

CENELEC EN 613326-1: 2006 Electrical Equipment for Measurement, Control and Laboratory Use – EMC Requirements Part - General Requirements

CISPR11: 2004 Industrial, Scientific and Medical (ISM) Radio-Frequency Equipment-Electromagnetic Disturbance Characteristics-Limits and Methods of Measurement

IEC 61000-4-2: 2001 Electromagnetic Compatibility (EMC) - Part 4-2: Test and Measurement Techniques – Electrostatic Discharge Immunity Test

IEC 61000-4-3: 2006 Electromagnetic Compatibility (EMC) - Part 4-3: Testing and Measurement Techniques – Radiated, Radio-Frequency, Electromagnetic Field Immunity Test

The authorized representative located within the Community is:

Dwyer Instruments Ltd  
Unit 16, The Wye Estate, London Road  
High Wycombe, Bucks HP11 1LH-U.K.  
(+44) (0) 1494 461707

On behalf of Dwyer Instruments, Inc.

Senior Regulatory Engineer

Michigan City, Indiana, USA

7/30/10